

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081919\
 Data File : VX011657.D
 Acq On : 20 Aug 2019 01:53
 Operator : JC/SP
 Sample : K4385-28
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-A3-(2)

Quant Time: Aug 20 13:43:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	416233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	539699	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	532138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	226681	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	211101	54.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.38%	
35) Dibromofluoromethane	5.49	113	175948	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	8.71	98	679843	61.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.62%	
62) 4-Bromofluorobenzene	11.13	95	237332	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
Target Compounds						
16) Acetone	2.43	43	8889	4.580	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081919\
Data File : VX011657.D
Acq On : 20 Aug 2019 01:53
Operator : JC/SP
Sample : K4385-28
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-A3-(2)

Quant Time: Aug 20 13:43:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

