

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081919\
 Data File : VX011665.D
 Acq On : 20 Aug 2019 05:01
 Operator : JC/SP
 Sample : K4385-40
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T8-E1-(0)

Quant Time: Aug 20 14:15:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	394098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	559856	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	508023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	252765	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	196976	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
35) Dibromofluoromethane	5.49	113	170733	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
50) Toluene-d8	8.71	98	655142	57.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.84%	
62) 4-Bromofluorobenzene	11.13	95	228641	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
Target Compounds						
16) Acetone	2.43	43	9240	5.028	ug/l	89
43) Isopropyl Acetate	6.44	43	56849	6.616	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081919\
Data File : VX011665.D
Acq On : 20 Aug 2019 05:01
Operator : JC/SP
Sample : K4385-40
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T8-E1-(0)

Quant Time: Aug 20 14:15:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

