

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX017972.D
 Acq On : 19 Aug 2020 10:15
 Operator : JC/SP
 Sample : VX0819WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0819WBL01

Quant Time: Aug 19 13:26:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	506650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	689190	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	689096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	276202	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	314607	42.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.72%	
35) Dibromofluoromethane	5.46	113	240711	46.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.80%	
50) Toluene-d8	8.70	98	814264	44.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.76%	
62) 4-Bromofluorobenzene	11.12	95	289054	43.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX017972.D
Acq On : 19 Aug 2020 10:15
Operator : JC/SP
Sample : VX0819WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0819WBL01

Quant Time: Aug 19 13:26:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

