

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX017973.D
 Acq On : 19 Aug 2020 10:46
 Operator : JC/SP
 Sample : L3711-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20200814

Quant Time: Aug 19 13:28:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	523466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	815745	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	742908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	349350	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	345293	45.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.00%	
35) Dibromofluoromethane	5.46	113	280507	46.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.36%	
50) Toluene-d8	8.70	98	964533	44.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.84%	
62) 4-Bromofluorobenzene	11.12	95	346162	43.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX017973.D
Acq On : 19 Aug 2020 10:46
Operator : JC/SP
Sample : L3711-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20200814

Quant Time: Aug 19 13:28:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

