

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX017974.D
 Acq On : 19 Aug 2020 11:09
 Operator : JC/SP
 Sample : L3717-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-TB-20200812

Quant Time: Aug 19 13:33:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	479264	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	765399	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	699885	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	319017	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	330001	46.97	ug/l	0.00
Spiked Amount			Recovery	=	93.94%	
35) Dibromofluoromethane	5.46	113	247018	43.34	ug/l	0.00
Spiked Amount			Recovery	=	86.68%	
50) Toluene-d8	8.70	98	824845	40.94	ug/l	0.00
Spiked Amount			Recovery	=	81.88%	
62) 4-Bromofluorobenzene	11.12	95	325907	43.80	ug/l	0.00
Spiked Amount			Recovery	=	87.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX017974.D
Acq On : 19 Aug 2020 11:09
Operator : JC/SP
Sample : L3717-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-TB-20200812

Quant Time: Aug 19 13:33:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

