

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX017976.D
 Acq On : 19 Aug 2020 11:54
 Operator : JC/SP
 Sample : L3717-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-GW-758-760

Quant Time: Aug 19 13:43:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	491003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	750112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	675273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	306373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	320773	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
35) Dibromofluoromethane	5.47	113	261270	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
50) Toluene-d8	8.70	98	905864	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
62) 4-Bromofluorobenzene	11.12	95	313967	43.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.10%	
Target Compounds						
16) Acetone	2.42	43	20165	7.121	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX017976.D
Acq On : 19 Aug 2020 11:54
Operator : JC/SP
Sample : L3717-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-GW-758-760

Quant Time: Aug 19 13:43:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

