

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX017988.D
 Acq On : 19 Aug 2020 16:27
 Operator : JC/SP
 Sample : L3721-01 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-OC-6

Quant Time: Aug 20 05:16:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	458340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	709257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	685303	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	338316	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	306604	45.63	ug/l	0.00
Spiked Amount			Recovery	=	91.26%	
35) Dibromofluoromethane	5.46	113	248765	47.10	ug/l	0.00
Spiked Amount			Recovery	=	94.20%	
50) Toluene-d8	8.70	98	864414	46.30	ug/l	0.00
Spiked Amount			Recovery	=	92.60%	
62) 4-Bromofluorobenzene	11.12	95	325810	47.25	ug/l	0.00
Spiked Amount			Recovery	=	94.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	142574	53.938	ug/l	95
17) Carbon Disulfide	2.54	76	5194128	518.189	ug/l	100
25) 2-Butanone	4.65	43	19843	4.959	ug/l	97
40) Benzene	6.11	78	48260	2.699	ug/l	100
52) Toluene	8.77	92	11723	1.002	ug/l	91
68) m/p-Xylenes	10.34	106	8515	0.986	ug/l	95
69) o-Xylene	10.68	106	3460	0.416	ug/l	94
84) 1,2,4-Trimethylbenzene	11.79	105	4920	0.249	ug/l	98
95) Naphthalene	13.82	128	1207125	57.266	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX017988.D
Acq On : 19 Aug 2020 16:27
Operator : JC/SP
Sample : L3721-01 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-OC-6

Quant Time: Aug 20 05:16:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

