

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX018007.D
 Acq On : 20 Aug 2020 01:12
 Operator : JC/SP
 Sample : L3694-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SLUICEWAY-081320

Quant Time: Aug 20 06:07:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	499987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	783050	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	745937	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	359010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	341708	46.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.24%	
35) Dibromofluoromethane	5.46	113	274586	47.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.18%	
50) Toluene-d8	8.70	98	962061	46.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.34%	
62) 4-Bromofluorobenzene	11.12	95	363565	47.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.52%	
Target Compounds						
16) Acetone	2.42	43	66666	23.120	ug/l	97
19) Methyl tert-butyl Ether	3.17	73	967118	56.130	ug/l	99
30) Chloroform	5.18	83	47364	4.639	ug/l	96
47) Bromodichloromethane	7.88	83	21785	2.542	ug/l	89
52) Toluene	8.76	92	4613	0.357	ug/l	98
60) Dibromochloromethane	9.56	129	7962	1.074	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX018007.D
Acq On : 20 Aug 2020 01:12
Operator : JC/SP
Sample : L3694-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SLUICEWAY-081320

Quant Time: Aug 20 06:07:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

