

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX018008.D
 Acq On : 20 Aug 2020 01:35
 Operator : JC/SP
 Sample : L3694-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMP-081320

Quant Time: Aug 20 06:09:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	485002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	748306	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	724349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	344419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	329067	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
35) Dibromofluoromethane	5.47	113	261187	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
50) Toluene-d8	8.70	98	917020	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
62) 4-Bromofluorobenzene	11.12	95	342842	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
Target Compounds						
16) Acetone	2.42	43	12361	4.419	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	9422	1.777	ug/l	90
44) Trichloroethene	7.20	130	3051	0.545	ug/l	83
64) Tetrachloroethene	9.32	164	128004	23.140	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX018008.D
Acq On : 20 Aug 2020 01:35
Operator : JC/SP
Sample : L3694-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMP-081320

Quant Time: Aug 20 06:09:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

