

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX018009.D
 Acq On : 20 Aug 2020 01:58
 Operator : JC/SP
 Sample : L3694-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP

Quant Time: Aug 20 06:11:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	482259	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	753755	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.10	117	703119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	309692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	333108	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
35) Dibromofluoromethane	5.47	113	263221	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
50) Toluene-d8	8.70	98	907439	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
62) 4-Bromofluorobenzene	11.12	95	327932	44.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.50%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.32	50	5413	0.821	ug/l	95
16) Acetone	2.42	43	10174	3.658	ug/l	90
20) Methylene Chloride	2.84	84	7652	1.493	ug/l #	91
27) cis-1,2-Dichloroethene	4.57	96	26344	4.997	ug/l	86
44) Trichloroethene	7.19	130	6915	1.227	ug/l	89
64) Tetrachloroethene	9.32	164	80005	14.900	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX018009.D
Acq On : 20 Aug 2020 01:58
Operator : JC/SP
Sample : L3694-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP

Quant Time: Aug 20 06:11:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

