

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081920\
 Data File : VX018014.D
 Acq On : 20 Aug 2020 03:52
 Operator : JC/SP
 Sample : L3738-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB181-GW-438-440

Quant Time: Aug 20 06:22:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	511459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	800871	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	747344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	373106	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	352985	47.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.16%	
35) Dibromofluoromethane	5.47	113	282313	47.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.68%	
50) Toluene-d8	8.70	98	964865	45.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.54%	
62) 4-Bromofluorobenzene	11.12	95	354345	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081920\
Data File : VX018014.D
Acq On : 20 Aug 2020 03:52
Operator : JC/SP
Sample : L3738-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB181-GW-438-440

Quant Time: Aug 20 06:22:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

