


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081920\
 Data File : VX017962.D
 Acq On : 18 Aug 2020 17:20
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Title : SW846 8260
 Last Update : Wed Aug 19 05:36:51 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1635

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	24936	PASS
75	95	30	60	50.7	72277	PASS
95	95	100	100	100.0	142514	PASS
96	95	5	9	6.5	9288	PASS
173	174	0.00	2	1.8	2017	PASS
174	95	50	100	80.6	114837	PASS
175	174	5	9	7.5	8656	PASS
176	174	95	101	95.5	109664	PASS
177	176	5	9	6.2	6822	PASS