


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081920\
Data File : VX017996.D
Acq On : 19 Aug 2020 19:53
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Title : SW846 8260
Last Update : Wed Aug 19 05:36:51 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	21574	PASS
75	95	30	60	50.3	58728	PASS
95	95	100	100	100.0	116712	PASS
96	95	5	9	6.2	7259	PASS
173	174	0.00	2	1.3	1254	PASS
174	95	50	100	83.4	97381	PASS
175	174	5	9	7.8	7581	PASS
176	174	95	101	96.1	93584	PASS
177	176	5	9	6.6	6165	PASS