

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
 Data File : VX047408.D
 Acq On : 19 Aug 2025 13:41
 Operator : JC/MD
 Sample : Q2815-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Aug 20 01:26:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	205616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	407837	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	405367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	203950	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	158172	54.965	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.940%
35) Dibromofluoromethane	5.403	113	134553	50.834	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.660%
50) Toluene-d8	8.653	98	470973	49.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.079	95	195434	55.979	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
Data File : VX047408.D
Acq On : 19 Aug 2025 13:41
Operator : JC/MD
Sample : Q2815-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Aug 20 01:26:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

