

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
 Data File : VX047417.D
 Acq On : 19 Aug 2025 16:57
 Operator : JC/MD
 Sample : Q2869-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-081325

Quant Time: Aug 20 01:29:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	205773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	414097	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	417583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	212656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	164219	57.023	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.040%	
35) Dibromofluoromethane	5.397	113	137119	51.021	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.040%	
50) Toluene-d8	8.653	98	483216	50.304	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.600%	
62) 4-Bromofluorobenzene	11.079	95	199041	56.150	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.300%	
Target Compounds						
16) Acetone	2.398	43	8905	8.226	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
Data File : VX047417.D
Acq On : 19 Aug 2025 16:57
Operator : JC/MD
Sample : Q2869-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-081325

Quant Time: Aug 20 01:29:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

