

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
 Data File : VX047419.D
 Acq On : 19 Aug 2025 17:39
 Operator : JC/MD
 Sample : Q2815-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-84SB-W

Quant Time: Aug 20 01:30:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	257962	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	524123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	526082	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	264563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	206988	57.333	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	114.660%	
35) Dibromofluoromethane	5.404	113	174696	51.357	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.720%	
50) Toluene-d8	8.653	98	621257	51.098	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.200%	
62) 4-Bromofluorobenzene	11.079	95	251572	56.071	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	112.140%	
Target Compounds						
16) Acetone	2.392	43	6455	4.757	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
Data File : VX047419.D
Acq On : 19 Aug 2025 17:39
Operator : JC/MD
Sample : Q2815-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-84SB-W

Quant Time: Aug 20 01:30:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

