

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
 Data File : VX047423.D
 Acq On : 19 Aug 2025 19:04
 Operator : JC/MD
 Sample : Q2844-09DL 10X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0226DL

Quant Time: Aug 20 01:32:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	350398	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	348764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	172723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	141737	57.649	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 115.300%	
35) Dibromofluoromethane	5.403	113	117945	51.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.720%	
50) Toluene-d8	8.653	98	418447	51.480	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.960%	
62) 4-Bromofluorobenzene	11.079	95	170785	56.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.880%	
Target Compounds						
					Qvalue	
52) Toluene	8.720	92	290195	43.706	ug/l	99
67) Ethyl Benzene	10.195	91	25914	1.817	ug/l	98
68) m/p-Xylenes	10.299	106	45040	8.468	ug/l	95
69) o-Xylene	10.640	106	14565	2.845	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081925\
Data File : VX047423.D
Acq On : 19 Aug 2025 19:04
Operator : JC/MD
Sample : Q2844-09DL 10X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0226DL

Quant Time: Aug 20 01:32:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

