

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082018\
 Data File : VX004150.D
 Acq On : 20 Aug 2018 14:03
 Operator : JC/MD
 Sample : J4541-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY070MW236-180815

Quant Time: Aug 21 11:21:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	286650	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	439770	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	416680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	218016	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
35) Dibromofluoromethane	5.51	113	188304	57.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.42%	
50) Toluene-d8	8.72	98	658495	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	11.14	95	218108	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	1962	0.57	ug/l	# 80
27) cis-1,2-Dichloroethene	4.60	96	18808	4.88	ug/l	87
44) Trichloroethene	7.21	130	1672	0.41	ug/l	96
64) Tetrachloroethene	9.34	164	1511	0.38	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082018\
Data File : VX004150.D
Acq On : 20 Aug 2018 14:03
Operator : JC/MD
Sample : J4541-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY070MW236-180815

Quant Time: Aug 21 11:21:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

