

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082019\
 Data File : VX011685.D
 Acq On : 20 Aug 2019 15:25
 Operator : JC/SP
 Sample : K4397-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 21 12:05:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	411578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	571073	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	501482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	241079	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	205772	53.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.82%	
35) Dibromofluoromethane	5.49	113	169971	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	
50) Toluene-d8	8.71	98	641125	55.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.18%	
62) 4-Bromofluorobenzene	11.13	95	223145	45.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082019\
Data File : VX011685.D
Acq On : 20 Aug 2019 15:25
Operator : JC/SP
Sample : K4397-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Aug 21 12:05:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

