

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082019\
 Data File : VX011699.D
 Acq On : 20 Aug 2019 21:40
 Operator : JC/SP
 Sample : PB122377TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB122377TB

Quant Time: Aug 21 00:44:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	351239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	518409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	441716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	234282	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	186629	57.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.60%	
35) Dibromofluoromethane	5.49	113	161163	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
50) Toluene-d8	8.71	98	626580	59.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.62%	
62) 4-Bromofluorobenzene	11.13	95	214032	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
Target Compounds						
20) Methylene Chloride	2.85	84	4378	1.273	ug/l	92
43) Isopropyl Acetate	6.44	43	64474	8.104	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082019\
Data File : VX011699.D
Acq On : 20 Aug 2019 21:40
Operator : JC/SP
Sample : PB122377TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB122377TB

Quant Time: Aug 21 00:44:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

