

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082019\  
 Data File : VX011715.D  
 Acq On : 21 Aug 2019 03:57  
 Operator : JC/SP  
 Sample : K4387-40  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 T4-B2-(0)

Quant Time: Aug 21 05:49:21 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X081519W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 16 04:15:26 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min)  |
|-----------------------------|-------|------|----------|-------|---------|-----------|
| 1) Pentafluorobenzene       | 5.66  | 168  | 360400   | 50.00 | ug/l    | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.85  | 114  | 511202   | 50.00 | ug/l    | 0.00      |
| 63) Chlorobenzene-d5        | 10.11 | 117  | 465352   | 50.00 | ug/l    | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.07 | 152  | 219037   | 50.00 | ug/l    | 0.00      |
| System Monitoring Compounds |       |      |          |       |         |           |
| 33) 1,2-Dichloroethane-d4   | 6.06  | 65   | 188653   | 56.45 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 112.90% |           |
| 35) Dibromofluoromethane    | 5.49  | 113  | 157546   | 47.12 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 94.24%  |           |
| 50) Toluene-d8              | 8.71  | 98   | 600819   | 57.67 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 115.34% |           |
| 62) 4-Bromofluorobenzene    | 11.13 | 95   | 195034   | 44.35 | ug/l    | 0.00      |
| Spiked Amount 50.000        |       |      | Recovery | =     | 88.70%  |           |
| Target Compounds            |       |      |          |       |         |           |
| 43) Isopropyl Acetate       | 6.45  | 43   | 55838    | 7.117 | ug/l    | Qvalue 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082019\  
Data File : VX011715.D  
Acq On : 21 Aug 2019 03:57  
Operator : JC/SP  
Sample : K4387-40  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 48 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
T4-B2-(0)

Quant Time: Aug 21 05:49:21 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X081519W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 16 04:15:26 2019  
Response via : Initial Calibration

