

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082019\
 Data File : VX011717.D
 Acq On : 21 Aug 2019 04:44
 Operator : JC/SP
 Sample : K4388-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T4-B4-(0)

Quant Time: Aug 21 05:54:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	349555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	495690	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	471052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	218466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	183614	56.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.28%	
35) Dibromofluoromethane	5.49	113	151503	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
50) Toluene-d8	8.71	98	575187	56.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.88%	
62) 4-Bromofluorobenzene	11.13	95	199398	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
Target Compounds						
16) Acetone	2.43	43	8674	5.322	ug/l	97
43) Isopropyl Acetate	6.43	43	52616	6.916	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082019\
Data File : VX011717.D
Acq On : 21 Aug 2019 04:44
Operator : JC/SP
Sample : K4388-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T4-B4-(0)

Quant Time: Aug 21 05:54:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

