

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082019\
 Data File : VX011676.D
 Acq On : 20 Aug 2019 11:54
 Operator : JC/SP
 Sample : K4406-02RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20190816RE

Quant Time: Aug 20 13:26:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	432578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	586600	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	522591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	258549	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	210206	52.40	ug/l	0.00
Spiked Amount						
			Recovery	=	104.80%	
35) Dibromofluoromethane	5.49	113	170647	44.48	ug/l	0.00
Spiked Amount						
			Recovery	=	88.96%	
50) Toluene-d8	8.71	98	677786	56.70	ug/l	0.00
Spiked Amount						
			Recovery	=	113.40%	
62) 4-Bromofluorobenzene	11.13	95	238090	47.18	ug/l	0.00
Spiked Amount						
			Recovery	=	94.36%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082019\
Data File : VX011676.D
Acq On : 20 Aug 2019 11:54
Operator : JC/SP
Sample : K4406-02RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190816RE

Quant Time: Aug 20 13:26:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 738

Delta R.T. 0.00 min

Lab File: VX011676.D

Acq: 20 Aug 2019 11:54

Instrument :

MSVOA_X

ClientSampleId :


TT-TB-20190816RE

Tot Ion:168 Resp: 432578

Ion Ratio Lower Upper

168 100

99 47.0 37.3 55.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 52.400 ug/l

RT: 6.06 min Scan# 804

Delta R.T. 0.00 min

Lab File: VX011676.D


Acq: 20 Aug 2019 11:54

Tgt Ion: 65 Resp: 210206

Ion Ratio Lower Upper

65 100

67 51.6 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. 0.00 min

Lab File: VX011676.D

Acq: 20 Aug 2019 11:54

Instrument :

MSVOA_X

ClientSampleId :

TT-TB-20190816RE


Tot Ion:114 Resp: 586600

Ion Ratio Lower Upper

114 100

63 18.8 0.0 37.8




88 14.7 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 44.477 ug/l

RT: 5.49 min Scan# 711

Delta R.T. -0.01 min

Lab File: VX011676.D

Acq: 20 Aug 2019 11:54


Tgt Ion:113 Resp: 170647

Ion Ratio Lower Upper

113 100

111 102.2 81.1 121.7



192 25.4 20.8 31.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 56.696 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX011676.D

Acq: 20 Aug 2019 11:54

Instrument :

MSVOA_X

ClientSampleId :





TT-TB-20190816RE

Tot Ion: 98 Resp: 677786

Ion Ratio Lower Upper

98 100

100 64.5 52.4 78.6



#62

4-Bromofluorobenzene

Concen: 47.185 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX011676.D

Acq: 20 Aug 2019 11:54




Tgt Ion: 95 Resp: 238090


Ion Ratio Lower Upper

95 100

174 95.6 0.0 200.6

176 93.3 0.0 196.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX011676.D
Acq: 20 Aug 2019 11:54

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20190816RE

Tot Ion:117 Resp: 522591
Ion Ratio Lower Upper
117 100
82 52.1 42.0 63.0
119 32.0 26.5 39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1791
Delta R.T. 0.00 min
Lab File: VX011676.D
Acq: 20 Aug 2019 11:54

Tgt Ion:152 Resp: 258549
Ion Ratio Lower Upper
152 100
115 53.9 37.3 111.9
150 157.9 0.0 346.2

