

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082020\
 Data File : VX018030.D
 Acq On : 20 Aug 2020 13:24
 Operator : JC/SP
 Sample : L3694-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKS

Quant Time: Aug 21 01:45:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	583282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	921355	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	851149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	417383	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	376633	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
35) Dibromofluoromethane	5.47	113	310213	45.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.42%	
50) Toluene-d8	8.70	98	1100237	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
62) 4-Bromofluorobenzene	11.12	95	412356	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
Target Compounds						
16) Acetone	2.42	43	69925	20.787	ug/l	99
20) Methylene Chloride	2.84	84	4643	0.749	ug/l #	82
30) Chloroform	5.18	83	233673	19.620	ug/l	99
47) Bromodichloromethane	7.88	83	66342	6.580	ug/l	91
60) Dibromochloromethane	9.57	129	14056	1.612	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082020\
Data File : VX018030.D
Acq On : 20 Aug 2020 13:24
Operator : JC/SP
Sample : L3694-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKS

Quant Time: Aug 21 01:45:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

