

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082020\
 Data File : VX018033.D
 Acq On : 20 Aug 2020 14:33
 Operator : JC/SP
 Sample : L3721-01DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 I-OC-6DL

Quant Time: Aug 21 02:01:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 05:36:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	580891	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	936145	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	887608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	446402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	382713	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
35) Dibromofluoromethane	5.47	113	310797	44.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.16%	
50) Toluene-d8	8.70	98	1137668	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
62) 4-Bromofluorobenzene	11.12	95	425783	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
Target Compounds						
16) Acetone	2.42	43	19229	5.740	ug/l	91
17) Carbon Disulfide	2.55	76	557882	43.915	ug/l	100
95) Naphthalene	13.82	128	147482	5.302	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082020\
Data File : VX018033.D
Acq On : 20 Aug 2020 14:33
Operator : JC/SP
Sample : L3721-01DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
I-OC-6DL

Quant Time: Aug 21 02:01:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081920W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 05:36:51 2020
Response via : Initial Calibration

