

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023846.D
 Acq On : 19 Aug 2021 18:34
 Operator : JC/MD
 Sample : M3483-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BB-TT-VPB-RW7-GW-245-247

Quant Time: Aug 20 03:15:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237112	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98485	50.874	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.740%
35) Dibromofluoromethane	5.397	113	77031	49.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	288742	48.340	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.085	95	96213	43.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.520%
Target Compounds						
16) Acetone	2.386	43	9343	10.794	ug/l	92
25) 2-Butanone	4.574	43	3405	2.597	ug/l #	75
30) Chloroform	5.105	83	3230	1.122	ug/l	96
44) Trichloroethene	7.129	130	2304	1.285	ug/l	72
64) Tetrachloroethene	9.275	164	740	0.424	ug/l #	73
95) Naphthalene	13.780	128	2162	1.619	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023846.D
Acq On : 19 Aug 2021 18:34
Operator : JC/MD
Sample : M3483-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BB-TT-VPB-RW7-GW-245-247

Quant Time: Aug 20 03:15:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

