

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023848.D
 Acq On : 19 Aug 2021 19:22
 Operator : JC/MD
 Sample : M3484-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Aug 20 03:15:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	220190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89334	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98071	51.555	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.120%
35) Dibromofluoromethane	5.397	113	77284	49.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	287514	48.191	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.380%
62) 4-Bromofluorobenzene	11.085	95	96665	44.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023848.D
Acq On : 19 Aug 2021 19:22
Operator : JC/MD
Sample : M3484-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Aug 20 03:15:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

