

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023852.D
 Acq On : 19 Aug 2021 20:56
 Operator : JC/MD
 Sample : M3483-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BB-TT-VPB-RW7-GW-280-282

Quant Time: Aug 20 03:16:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	134862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	97586	51.845	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.700%			
35) Dibromofluoromethane	5.397	113	77419	49.968	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.940%			
50) Toluene-d8	8.653	98	287876	48.817	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.640%			
62) 4-Bromofluorobenzene	11.085	95	96422	44.417	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 88.840%			
Target Compounds						Qvalue
16) Acetone	2.386	43	3569	4.241	ug/l #	82
17) Carbon Disulfide	2.514	76	2246	0.728	ug/l	97
19) Methyl tert-butyl Ether	3.130	73	1707	0.379	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023852.D
Acq On : 19 Aug 2021 20:56
Operator : JC/MD
Sample : M3483-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BB-TT-VPB-RW7-GW-280-282

Quant Time: Aug 20 03:16:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

