

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023858.D
 Acq On : 19 Aug 2021 23:17
 Operator : JC/MD
 Sample : M3484-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 O22-OUTFALL

Quant Time: Aug 20 03:17:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132408	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223313	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95402	51.625	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 103.240%			
35) Dibromofluoromethane	5.397	113	74619	50.485	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.960%			
50) Toluene-d8	8.653	98	279864	49.749	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.500%			
62) 4-Bromofluorobenzene	11.085	95	96253	46.478	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.960%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	70623	85.470	ug/l	90
25) 2-Butanone	4.574	43	20770	16.596	ug/l	93
51) 4-Methyl-2-Pentanone	8.580	43	7288	2.835	ug/l #	85
70) Styrene	10.659	104	14311	3.007	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023858.D
Acq On : 19 Aug 2021 23:17
Operator : JC/MD
Sample : M3484-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
O22-OUTFALL

Quant Time: Aug 20 03:17:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

