

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023860.D
 Acq On : 20 Aug 2021 00:04
 Operator : JC/MD
 Sample : M3468-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SAMPLE-9-(322.00)

Quant Time: Aug 20 03:17:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227293	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96033	52.124	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.240%
35) Dibromofluoromethane	5.397	113	74514	49.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	8.653	98	284543	49.695	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.085	95	92893	44.071	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.140%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	23689	28.756	ug/l #	87
20) Methylene Chloride	2.794	84	12855	8.169	ug/l #	87
43) Isopropyl Acetate	6.355	43	16567	4.093	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023860.D
Acq On : 20 Aug 2021 00:04
Operator : JC/MD
Sample : M3468-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SAMPLE-9-(322.00)

Quant Time: Aug 20 03:17:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

