

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
 Data File : VX023861.D
 Acq On : 20 Aug 2021 00:28
 Operator : JC/MD
 Sample : M3468-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SAMPLE-8-(309.00)

Quant Time: Aug 20 03:17:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	136190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232846	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97741	51.421	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.840%	
35) Dibromofluoromethane	5.397	113	76220	49.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.920%	
50) Toluene-d8	8.653	98	295556	50.387	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.780%	
62) 4-Bromofluorobenzene	11.085	95	99207	45.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.880%	
Target Compounds						Qvalue
16) Acetone	2.386	43	21021	24.734	ug/l	88
20) Methylene Chloride	2.794	84	11289	6.954	ug/l	91
43) Isopropyl Acetate	6.355	43	14493	3.496	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082021\
Data File : VX023861.D
Acq On : 20 Aug 2021 00:28
Operator : JC/MD
Sample : M3468-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SAMPLE-8-(309.00)

Quant Time: Aug 20 03:17:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

