

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047432.D
 Acq On : 20 Aug 2025 11:28
 Operator : JC/MD
 Sample : Q2815-21RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-11M-ERE

Quant Time: Aug 21 01:28:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	288810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	596562	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	600326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	298812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	240940	59.609	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 119.220%#		
35) Dibromofluoromethane	5.397	113	204272	52.760	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.520%		
50) Toluene-d8	8.653	98	718573	51.925	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.860%		
62) 4-Bromofluorobenzene	11.079	95	298984	58.547	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 117.100%		
Target Compounds						
16) Acetone	2.392	43	24055	15.833	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047432.D
Acq On : 20 Aug 2025 11:28
Operator : JC/MD
Sample : Q2815-21RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-11M-ERE

Quant Time: Aug 21 01:28:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

