

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047433.D
 Acq On : 20 Aug 2025 11:49
 Operator : JC/MD
 Sample : Q2815-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-11M-S

Quant Time: Aug 21 01:29:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	275255	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	538445	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	528308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	264520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	211576	54.922	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.840%	
35) Dibromofluoromethane	5.404	113	178714	51.141	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.280%	
50) Toluene-d8	8.653	98	641607	51.368	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.740%	
62) 4-Bromofluorobenzene	11.079	95	256712	55.695	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.400%	
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	7219	4.985	ug/l	96
17) Carbon Disulfide	2.532	76	8056	0.770	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047433.D
Acq On : 20 Aug 2025 11:49
Operator : JC/MD
Sample : Q2815-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-11M-S

Quant Time: Aug 21 01:29:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

