

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047434.D
 Acq On : 20 Aug 2025 12:11
 Operator : JC/MD
 Sample : Q2815-26
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Aug 21 01:29:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	317644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	633243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	619750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	313519	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	249104	56.035	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.060%
35) Dibromofluoromethane	5.403	113	208667	50.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	8.653	98	748732	50.970	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	305594	56.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047434.D
Acq On : 20 Aug 2025 12:11
Operator : JC/MD
Sample : Q2815-26
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Aug 21 01:29:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

