

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047437.D
 Acq On : 20 Aug 2025 13:15
 Operator : JC/MD
 Sample : Q2869-03 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-01-6.5-081325

Quant Time: Aug 21 01:30:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	296536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	599838	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	598049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	298160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	242493	58.430	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.860%	
35) Dibromofluoromethane	5.397	113	205293	52.734	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.460%	
50) Toluene-d8	8.653	98	711871	51.160	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.320%	
62) 4-Bromofluorobenzene	11.079	95	291371	56.745	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.480%	
Target Compounds						Qvalue
16) Acetone	2.404	43	1536	0.985	ug/l	90
44) Trichloroethene	7.147	130	2156	0.458	ug/l	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047437.D
Acq On : 20 Aug 2025 13:15
Operator : JC/MD
Sample : Q2869-03 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-01-6.5-081325

Quant Time: Aug 21 01:30:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

