

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047439.D
 Acq On : 20 Aug 2025 13:59
 Operator : JC/MD
 Sample : Q2869-05 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19B-71.5-081325-FD

Quant Time: Aug 21 01:31:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	303425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	608029	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	599147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	301736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	237099	55.834	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.660%
35) Dibromofluoromethane	5.403	113	199244	50.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.980%
50) Toluene-d8	8.653	98	718098	50.912	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.820%
62) 4-Bromofluorobenzene	11.079	95	286076	54.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.920%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.355	101	12596	3.361	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	151640	32.109	ug/l	91
44) Trichloroethene	7.135	130	158224	33.141	ug/l	96
64) Tetrachloroethene	9.281	164	6917	1.596	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047439.D
Acq On : 20 Aug 2025 13:59
Operator : JC/MD
Sample : Q2869-05 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19B-71.5-081325-FD

Quant Time: Aug 21 01:31:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

