

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
 Data File : VX047440.D
 Acq On : 20 Aug 2025 14:20
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 21 01:31:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 04:18:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	295829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	604962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	597909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	298924	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	242154	58.488	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.980%
35) Dibromofluoromethane	5.404	113	203918	51.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.880%
50) Toluene-d8	8.653	98	722220	51.464	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.920%
62) 4-Bromofluorobenzene	11.079	95	291944	56.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082025\
Data File : VX047440.D
Acq On : 20 Aug 2025 14:20
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 21 01:31:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 04:18:28 2025
Response via : Initial Calibration

