

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004169.D
 Acq On : 21 Aug 2018 12:19
 Operator : JC/MD
 Sample : J4560-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-0818-S-TCLP-GRID61-2(2)

Quant Time: Aug 22 01:30:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	284055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	454750	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	442272	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	226562	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	202785	50.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.34%	
35) Dibromofluoromethane	5.50	113	174445	51.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.52%	
50) Toluene-d8	8.71	98	646470	46.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.70%	
62) 4-Bromofluorobenzene	11.14	95	213285	42.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.24%	

Target Compounds						Qvalue
10) Methyl Iodide	2.51	142	1331	5.36	ug/l	99
16) Acetone	2.44	43	23836	11.46	ug/l	98
18) Methyl Acetate	2.78	43	6598	1.43	ug/l	99
20) Methylene Chloride	2.86	84	16273	4.23	ug/l	97
25) 2-Butanone	4.71	43	4481	1.32	ug/l	95
43) Isopropyl Acetate	6.46	43	152258	18.33	ug/l	100
95) Naphthalene	13.83	128	1448	1.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004169.D
Acq On : 21 Aug 2018 12:19
Operator : JC/MD
Sample : J4560-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-0818-S-TCLP-GRID61-2(2)

Quant Time: Aug 22 01:30:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

