

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004171.D
 Acq On : 21 Aug 2018 13:12
 Operator : JC/MD
 Sample : J4545-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB-BASELINE-WATER

Quant Time: Aug 22 01:23:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	306702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	508011	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	490179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	270640	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	217919	50.43	ug/l	0.00
Spiked Amount			Recovery	=	100.86%	
35) Dibromofluoromethane	5.50	113	191883	50.47	ug/l	0.00
Spiked Amount			Recovery	=	100.94%	
50) Toluene-d8	8.71	98	724963	47.03	ug/l	0.00
Spiked Amount			Recovery	=	94.06%	
62) 4-Bromofluorobenzene	11.14	95	271056	48.48	ug/l	0.00
Spiked Amount			Recovery	=	96.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	4395	1.27	ug/l	95
10) Methyl Iodide	2.51	142	1147	5.29	ug/l	# 87
11) Tert butyl alcohol	3.07	59	158098	181.68	ug/l	99
16) Acetone	2.44	43	452490	201.40	ug/l	99
25) 2-Butanone	4.70	43	197025	53.90	ug/l	99
40) Benzene	6.14	78	60384	3.48	ug/l	97
51) 4-Methyl-2-Pentanone	8.65	43	61947	8.14	ug/l	99
59) 2-Hexanone	9.50	43	56194	9.81	ug/l	85
83) tert-Butylbenzene	11.77	119	30282	2.00	ug/l	91
84) 1,2,4-Trimethylbenzene	11.81	105	26389	1.70	ug/l	98
95) Naphthalene	13.83	128	110804	6.87	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004171.D
Acq On : 21 Aug 2018 13:12
Operator : JC/MD
Sample : J4545-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB-BASELINE-WATER

Quant Time: Aug 22 01:23:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

