

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004173.D
 Acq On : 21 Aug 2018 14:05
 Operator : JC/MD
 Sample : J4271-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-06-082018-B

Quant Time: Aug 22 01:28:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	276658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	469726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	445551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	239652	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	207327	53.19	ug/l	0.00
Spiked Amount						
			Recovery	=	106.38%	
35) Dibromofluoromethane	5.50	113	176576	50.23	ug/l	0.00
Spiked Amount						
			Recovery	=	100.46%	
50) Toluene-d8	8.71	98	665920	46.72	ug/l	0.00
Spiked Amount						
			Recovery	=	93.44%	
62) 4-Bromofluorobenzene	11.14	95	227673	44.04	ug/l	0.00
Spiked Amount						
			Recovery	=	88.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	17109	8.44	ug/l	99
18) Methyl Acetate	2.78	43	7438	1.65	ug/l	99
20) Methylene Chloride	2.86	84	149726	45.14	ug/l	99
25) 2-Butanone	4.71	43	4399	1.33	ug/l #	82
30) Chloroform	5.21	83	13836	2.31	ug/l	97
43) Isopropyl Acetate	6.46	43	182506	21.27	ug/l	99
95) Naphthalene	13.83	128	221730	14.07	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004173.D
Acq On : 21 Aug 2018 14:05
Operator : JC/MD
Sample : J4271-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-06-082018-B

Quant Time: Aug 22 01:28:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

