

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004180.D
 Acq On : 21 Aug 2018 17:10
 Operator : JC/MD
 Sample : J4275-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-082018-A

Quant Time: Aug 22 01:47:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	296874	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	485204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	468512	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	245443	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	213552	51.05	ug/l	0.00
Spiked Amount			Recovery	=	102.10%	
35) Dibromofluoromethane	5.50	113	182757	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
50) Toluene-d8	8.71	98	688265	46.75	ug/l	0.00
Spiked Amount			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	11.14	95	244206	45.73	ug/l	0.00
Spiked Amount			Recovery	=	91.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.07	59	2282	2.71	ug/l	# 91
16) Acetone	2.45	43	25167	11.57	ug/l	100
18) Methyl Acetate	2.78	43	7551	1.56	ug/l	98
20) Methylene Chloride	2.85	84	8789	1.89	ug/l	96
25) 2-Butanone	4.70	43	4563	1.29	ug/l	94
43) Isopropyl Acetate	6.46	43	187537	21.16	ug/l	100
95) Naphthalene	13.83	128	1378792	79.65	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004180.D
Acq On : 21 Aug 2018 17:10
Operator : JC/MD
Sample : J4275-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-082018-A

Quant Time: Aug 22 01:47:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

