

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004181.D
 Acq On : 21 Aug 2018 17:37
 Operator : JC/MD
 Sample : J4275-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-082018-B

Quant Time: Aug 22 01:49:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	290076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	474424	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	455089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	244867	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	208483	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	5.50	113	177721	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
50) Toluene-d8	8.71	98	671251	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
62) 4-Bromofluorobenzene	11.14	95	229226	43.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.80%	
Target Compounds						
16) Acetone	2.44	43	22154	10.43	ug/l	95
18) Methyl Acetate	2.78	43	7612	1.61	ug/l	93
20) Methylene Chloride	2.86	84	2050217	596.95	ug/l	97
25) 2-Butanone	4.70	43	3967	1.15	ug/l	94
30) Chloroform	5.21	83	9086	1.44	ug/l	99
43) Isopropyl Acetate	6.46	43	180328	20.81	ug/l	100
95) Naphthalene	13.83	128	18626	2.21	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004181.D
Acq On : 21 Aug 2018 17:37
Operator : JC/MD
Sample : J4275-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-082018-B

Quant Time: Aug 22 01:49:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

