

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082118\
 Data File : VX004183.D
 Acq On : 21 Aug 2018 18:30
 Operator : JC/MD
 Sample : J4275-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-082018-D

Quant Time: Aug 22 01:54:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	449992	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	428746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	224907	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	200628	52.24	ug/l	0.00
Spiked Amount			Recovery	=	104.48%	
35) Dibromofluoromethane	5.51	113	171038	50.79	ug/l	0.00
Spiked Amount			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	644933	47.23	ug/l	0.00
Spiked Amount			Recovery	=	94.46%	
62) 4-Bromofluorobenzene	11.14	95	217797	43.98	ug/l	0.00
Spiked Amount			Recovery	=	87.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	19048	9.54	ug/l	99
18) Methyl Acetate	2.78	43	9004	2.03	ug/l	98
20) Methylene Chloride	2.85	84	7761606	2406.91	ug/l	95
25) 2-Butanone	4.70	43	4461	1.37	ug/l #	88
43) Isopropyl Acetate	6.46	43	183099	22.28	ug/l	100
95) Naphthalene	13.83	128	25289	2.72	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082118\
Data File : VX004183.D
Acq On : 21 Aug 2018 18:30
Operator : JC/MD
Sample : J4275-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-082018-D

Quant Time: Aug 22 01:54:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

