

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082119\
 Data File : VX011724.D
 Acq On : 21 Aug 2019 11:59
 Operator : JC/SP
 Sample : K4391-09ME
 Misc : 5.03g/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BOTTOM-2ME

Quant Time: Aug 21 13:58:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	423387	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.85	114	608617	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	554331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	308176	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	213847	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	5.49	113	165310	41.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.06%	
50) Toluene-d8	8.71	98	698492	56.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.62%	
62) 4-Bromofluorobenzene	11.13	95	257665	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
Target Compounds						
39) Methylcyclohexane	7.45	83	11019	1.715	ug/l	94
80) 1,3,5-Trimethylbenzene	11.51	105	22030	1.311	ug/l	96
84) 1,2,4-Trimethylbenzene	11.80	105	51251	3.039	ug/l	99
95) Naphthalene	13.83	128	65046	3.008	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082119\
Data File : VX011724.D
Acq On : 21 Aug 2019 11:59
Operator : JC/SP
Sample : K4391-09ME
Misc : 5.03g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BOTTOM-2ME

Quant Time: Aug 21 13:58:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

