

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082119\
 Data File : VX011727.D
 Acq On : 21 Aug 2019 13:10
 Operator : JC/SP
 Sample : K4413-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Aug 22 03:49:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	411039	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	631849	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	572159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	265152	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	223366	58.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.20%	
35) Dibromofluoromethane	5.49	113	179095	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
50) Toluene-d8	8.71	98	727182	56.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.94%	
62) 4-Bromofluorobenzene	11.13	95	240871	44.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.64%	
Target Compounds						
16) Acetone	2.43	43	8464	4.416	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082119\
Data File : VX011727.D
Acq On : 21 Aug 2019 13:10
Operator : JC/SP
Sample : K4413-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Aug 22 03:49:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

