

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082119\
 Data File : VX011745.D
 Acq On : 22 Aug 2019 01:29
 Operator : JC/SP
 Sample : K4096-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-081919

Quant Time: Aug 22 04:24:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	404007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	582924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	543016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	285924	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	220000	58.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.44%	
35) Dibromofluoromethane	5.49	113	184555	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	8.71	98	664303	55.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.84%	
62) 4-Bromofluorobenzene	11.13	95	251896	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
Target Compounds						
16) Acetone	2.43	43	24298	12.898	ug/l	98
20) Methylene Chloride	2.85	84	23054	5.828	ug/l	94
43) Isopropyl Acetate	6.43	43	72813	8.139	ug/l	98
95) Naphthalene	13.83	128	83521	4.163	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082119\
Data File : VX011745.D
Acq On : 22 Aug 2019 01:29
Operator : JC/SP
Sample : K4096-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-05-081919

Quant Time: Aug 22 04:24:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

