

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082120\
 Data File : VX018050.D
 Acq On : 21 Aug 2020 11:36
 Operator : JC/SP
 Sample : VX0821WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0821WBL01

Quant Time: Aug 22 00:01:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	651727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1073239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	993288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	463700	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	454159	50.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.40%	
35) Dibromofluoromethane	5.46	113	344371	47.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.22%	
50) Toluene-d8	8.70	98	1319607	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	
62) 4-Bromofluorobenzene	11.12	95	464340	43.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.70%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082120\
Data File : VX018050.D
Acq On : 21 Aug 2020 11:36
Operator : JC/SP
Sample : VX0821WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0821WBL01

Quant Time: Aug 22 00:01:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX018050.D

Acq: 21 Aug 2020 11:36

Instrument :

MSVOA_X

ClientSampleId :


VX0821WBL01

Tot Ion:168 Resp: 651727

Ion Ratio Lower Upper

168 100

99 57.1 46.3 69.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

250000

200000

150000

100000

50000



0

Time-->

5.50

5.60

5.70



#33

1,2-Dichloroethane-d4

Concen: 50.196 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX018050.D


Acq: 21 Aug 2020 11:36

Tgt Ion: 65 Resp: 454159

Ion Ratio Lower Upper

65 100

67 51.9 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.03

200000

150000

100000

50000


0


Time-->

5.90

6.00

6.10





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.01 min

Lab File: VX018050.D

Acq: 21 Aug 2020 11:36

Instrument :

MSVOA_X

ClientSampleId :

VX0821WBL01


Tot Ion:114 Resp: 1073239

Ion Ratio Lower Upper

114 100

63 21.5 0.0 42.6




88 16.8 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.114 ug/l

RT: 5.46 min Scan# 718

Delta R.T. -0.01 min

Lab File: VX018050.D

Acq: 21 Aug 2020 11:36


Tgt Ion:113 Resp: 344371

Ion Ratio Lower Upper

113 100

111 103.2 81.6 122.4



192 20.1 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.473 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX018050.D

Acq: 21 Aug 2020 11:36

Instrument :

MSVOA_X

ClientSampleId :





VX0821WBL01

Tot Ion: 98 Resp: 1319607

Ion Ratio Lower Upper

98 100

100 64.5 52.2 78.4



#62

4-Bromofluorobenzene

Concen: 43.352 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX018050.D

Acq: 21 Aug 2020 11:36




Tgt Ion: 95 Resp: 464340


Ion Ratio Lower Upper

95 100

174 77.7 0.0 160.2

176 73.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX018050.D
Acq: 21 Aug 2020 11:36

Instrument :
MSVOA_X
ClientSampleId :
VX0821WBL01

Tot Ion	Ratio	Lower	Upper
117	100		
82	54.7	44.2	66.2
119	32.8	25.7	38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX018050.D
Acq: 21 Aug 2020 11:36

Tgt Ion	Ratio	Lower	Upper
152	100		
115	59.2	41.5	124.5
150	156.8	0.0	343.8

