

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
 Data File : VX037130.D
 Acq On : 21 Aug 2023 14:55
 Operator : JC/MD
 Sample : 04086-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20230818

Quant Time: Aug 22 06:55:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193805	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102169	84.018	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 168.040%#	
35) Dibromofluoromethane	5.391	113	74993	58.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.860%	
50) Toluene-d8	8.647	98	237269	50.962	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	11.079	95	100246	54.789	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.580%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1470	3.090	ug/l	95
52) Toluene	8.720	92	9899	2.865	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
Data File : VX037130.D
Acq On : 21 Aug 2023 14:55
Operator : JC/MD
Sample : 04086-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20230818

Quant Time: Aug 22 06:55:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

