

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
 Data File : VX037134.D
 Acq On : 21 Aug 2023 16:32
 Operator : JC/MD
 Sample : 04086-13ME
 Misc : 4.99g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-PP17AME

Quant Time: Aug 22 05:58:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	144150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	238483	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	112277	54.027	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.060%	
35) Dibromofluoromethane	5.391	113	84740	53.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.300%	
50) Toluene-d8	8.646	98	329537	57.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 115.040%#	
62) 4-Bromofluorobenzene	11.085	95	130857	58.121	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
Data File : VX037134.D
Acq On : 21 Aug 2023 16:32
Operator : JC/MD
Sample : 04086-13ME
Misc : 4.99g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-PP17AME

Quant Time: Aug 22 05:58:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

