

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
 Data File : VX037135.D
 Acq On : 21 Aug 2023 16:56
 Operator : JC/MD
 Sample : 04086-15ME
 Misc : 5.00g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-PP18AME

Quant Time: Aug 22 05:58:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103516	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	116718	62.361	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	124.720%#
35) Dibromofluoromethane	5.391	113	87738	61.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	122.260%
50) Toluene-d8	8.647	98	339545	65.217	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	130.440%#
62) 4-Bromofluorobenzene	11.079	95	136104	66.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	133.040%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082123\
Data File : VX037135.D
Acq On : 21 Aug 2023 16:56
Operator : JC/MD
Sample : 04086-15ME
Misc : 5.00g/5.0mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-PP18AME

Quant Time: Aug 22 05:58:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

